

Only analyzers that have a passing Precision Calibration for the current quarter will show

PRECISION CALIBRATION LOG				
Analyzer: phx42-6252 ----- Calibration Quarter: 1Q2023				
FID: X PID:				
PPM	PRECAL BY	DATE	Avg Resp	Precision
95	S. Mortimer	1/10/2023	3.56	3.56
500	S. Mortimer	1/10/2023	2.48	3.52
10000	S. Mortimer	1/10/2023	2.26	1.76
TYPE OF GAS: Methane				