



LDARtools VWa Drift Report

Date Created: 1/31/2023

Analyzer ID	Date	Calibrate	Cyl. Mfr. ID	Conc.	Reading	Conc. / Reading Diff.	Probe Type	Pass	Drift 1	Drift1 Combined	Drift Pct1	Pass	Drift 2	Drift 2 Combined	Drift Pct2	Pass	Cal Note
phx42-8696	1/31/23	12:16:47 PM	CT-24077	0.00	1.73	17.30 %	Standard	Yes									
phx42-8696	1/31/23	12:17:24 PM	D3231001	95.00	96.40	1.47 %	Standard	Yes	01:27:38 PM	94.20 93.93 93.50	-1.18	Yes	01:43:46 PM	99.24 97.09 96.01	2.58	Yes	
phx42-8696	1/31/23	12:17:24 PM	D3231001	95.00	96.40	1.47 %	Standard	Yes	01:27:38 PM	94.20 93.93 93.50	-1.18	Yes	02:32:58 PM	97.30 95.60 95.07	1.04	Yes	
phx42-8696	1/31/23	12:17:24 PM	D3231001	95.00	96.40	1.47 %	Standard	Yes	02:03:29 PM	99.46 97.24 95.73	2.61	Yes	01:43:46 PM	99.24 97.09 96.01	2.58	Yes	
phx42-8696	1/31/23	12:17:24 PM	D3231001	95.00	96.40	1.47 %	Standard	Yes	02:03:29 PM	99.46 97.24 95.73	2.61	Yes	02:32:58 PM	97.30 95.60 95.07	1.04	Yes	
phx42-8696	1/31/23	12:18:00 PM	T7790071	492.00	487.00	-1.02 %	Standard	Yes	01:29:17 PM	484.00 484.00 485.00	-1.56	Yes	01:45:41 PM	480.00 485.00 486.00	-1.69	Yes	
phx42-8696	1/31/23	12:18:00 PM	T7790071	492.00	487.00	-1.02 %	Standard	Yes	01:29:17 PM	484.00 484.00 485.00	-1.56	Yes	02:34:37 PM	473.00 482.00 482.00	-2.64	Yes	
phx42-8696	1/31/23	12:18:00 PM	T7790071	492.00	487.00	-1.02 %	Standard	Yes	02:05:16 PM	479.00 485.00 485.00	-1.83	Yes	01:45:41 PM	480.00 485.00 486.00	-1.69	Yes	
phx42-8696	1/31/23	12:18:00 PM	T7790071	492.00	487.00	-1.02 %	Standard	Yes	02:05:16 PM	479.00 485.00 485.00	-1.83	Yes	02:34:37 PM	473.00 482.00 482.00	-2.64	Yes	
phx42-8696	1/31/23	12:18:36 PM	CT20546	9923.00	9809.00	-1.15 %	Standard	Yes	01:30:58 PM	9736.00 9746.00 9755.00	-1.79	Yes	01:47:21 PM	9663.00 9735.00 9722.00	-2.18	Yes	
phx42-8696	1/31/23	12:18:36 PM	CT20546	9923.00	9809.00	-1.15 %	Standard	Yes	01:30:58 PM	9736.00 9746.00 9755.00	-1.79	Yes	02:36:20 PM	9554.00 9683.00 9712.00	-2.75	Yes	
phx42-8696	1/31/23	12:18:36 PM	CT20546	9923.00	9809.00	-1.15 %	Standard	Yes	02:06:59 PM	9635.00 9724.00 9753.00	-2.21	Yes	01:47:21 PM	9663.00 9735.00 9722.00	-2.18	Yes	
phx42-8696	1/31/23	12:18:36 PM	CT20546	9923.00	9809.00	-1.15 %	Standard	Yes	02:06:59 PM	9635.00 9724.00 9753.00	-2.21	Yes	02:36:20 PM	9554.00 9683.00 9712.00	-2.75	Yes	

Zero gas was used to purge after the initial calibration process and between all samples during the drift assessment and precision calibration processes.