



# Analyzer Drift Report

Report run at: 9/29/2022

Analyzer	Date	Tech Id	Calibrate	Cylinder	Cyl. Exp	Conc.	Reading	Pass	Drift1	Drift1 1	Pct.	Pass	Drift2	Drift2 1	Pct.	Pass	Ext Probe
phx42-3536	9/22/2022	A. Scarborough	10:23:15 AM	CT-24077	7/18/2027	0	0.93	Yes									No
phx42-3536	9/22/2022	A. Scarborough	10:23:49 AM	CT8327	1/19/2027	98	98.77	Yes	01:41:01 PM	102.00	3.98 %	Yes	04:36:59 PM	98.40	0.31 %	Yes	No
phx42-3536	9/22/2022	A. Scarborough	10:24:23 AM	T7790071	1/18/2027	492	484.00	Yes	01:42:00 PM	501.00	1.83 %	Yes	04:37:28 PM	481.00	-2.24 %	Yes	No
phx42-3536	9/22/2022	A. Scarborough	10:24:57 AM	CT20546	10/6/2026	9923	9812.00	Yes	01:42:31 PM	9911.00	-0.12 %	Yes	04:37:58 PM	9883.00	-0.40 %	Yes	No
phx42-3584	9/22/2022	A. Scarborough	09:40:03 AM	CT-24077	7/18/2027	0	1.31	Yes									No
phx42-3584	9/22/2022	A. Scarborough	09:41:45 AM	CT8327	1/19/2027	98	97.99	Yes	01:27:27 PM	94.34	-3.83 %	Yes	04:04:56 PM	94.09	-4.09 %	Yes	No
phx42-3584	9/22/2022	A. Scarborough	09:43:26 AM	T7790071	1/18/2027	492	489.00	Yes	01:28:39 PM	480.00	-2.44 %	Yes	04:05:56 PM	474.00	-3.66 %	Yes	No
phx42-3584	9/22/2022	A. Scarborough	09:45:07 AM	CT20546	10/6/2026	9923	9808.00	Yes	01:29:44 PM	9599.00	-3.27 %	Yes	04:06:54 PM	9511.00	-4.15 %	Yes	No
phx42-3590	9/22/2022	A. Scarborough	09:39:29 AM	CT-24077	7/18/2027	0	0.82	Yes									No
phx42-3590	9/22/2022	A. Scarborough	09:41:12 AM	CT8327	1/19/2027	98	97.73	Yes	01:27:12 PM	93.52	-4.67 %	Yes	04:04:43 PM	95.60	-2.55 %	Yes	No
phx42-3590	9/22/2022	A. Scarborough	09:42:52 AM	T7790071	1/18/2027	492	488.00	Yes	01:28:19 PM	479.00	-2.64 %	Yes	04:05:44 PM	479.00	-2.64 %	Yes	No
phx42-3590	9/22/2022	A. Scarborough	09:44:34 AM	CT20546	10/6/2026	9923	9878.00	Yes	01:29:25 PM	9646.00	-2.79 %	Yes	04:06:43 PM	9584.00	-3.42 %	Yes	No
phx42-3599	9/22/2022	A. Scarborough	12:21:06 PM	CT-24077	7/18/2027	0	1.20	Yes									No
phx42-3599	9/22/2022	A. Scarborough	12:21:41 PM	CT8327	1/19/2027	98	98.11	Yes	04:05:03 PM	96.69	-1.44 %	Yes	06:40:40 PM	96.64	-1.49 %	Yes	No
phx42-3599	9/22/2022	A. Scarborough	12:22:15 PM	T7790071	1/18/2027	492	491.00	Yes	04:05:36 PM	478.00	-2.85 %	Yes	06:41:15 PM	475.00	-3.46 %	Yes	No
phx42-3599	9/22/2022	A. Scarborough	12:22:49 PM	CT20546	10/6/2026	9923	9804.00	Yes	04:06:18 PM	9761.00	-1.63 %	Yes	06:41:57 PM	9779.00	-1.45 %	Yes	No
phx42-5389	9/22/2022	S. Mort	10:32:11 AM	CT-24077	7/18/2027	0	1.21	Yes									No
phx42-5389	9/22/2022	S. Mort	10:32:45 AM	CT8327	1/19/2027	98	99.23	Yes	01:38:23 PM	102.00	3.98 %	Yes	04:33:11 PM	101.00	2.96 %	Yes	No
phx42-5389	9/22/2022	S. Mort	10:33:19 AM	T7790071	1/18/2027	492	486.00	Yes	01:39:29 PM	485.00	-1.42 %	Yes	04:33:42 PM	473.00	-3.86 %	Yes	No
phx42-5389	9/22/2022	S. Mort	10:33:53 AM	CT20546	10/6/2026	9923	9834.00	Yes	01:39:58 PM	9657.00	-2.68 %	Yes	04:34:13 PM	9478.00	-4.48 %	Yes	No

Zero gas was used to purge after the initial calibration process and between all samples during the drift assessment and precision calibration processes.