



LDARtools Chateau Reporting Daily Calibration

Date Created: 9/29/2022

Analyzer ID	Date	Tech Id	Probe Type	Probe Flow	Warm Up	Calibrate	Warm Up Duration	Cylinder	Cyl. Exp.	Conc	Reading	Conc. / Reading Diff.	Pass
phx42-3536	9/22/2022	A. Scarborough	Standard	0.30	08:59:25 AM	10:23:15 AM	84 min	CT-24077	7/18/2027	0.00	0.93	9.30 %	Yes
phx42-3536	9/22/2022	A. Scarborough	Standard	0.30	08:59:25 AM	10:23:49 AM	84 min	CT8327	1/19/2027	98.10	98.77	-0.68 %	Yes
phx42-3536	9/22/2022	A. Scarborough	Standard	0.30	08:59:25 AM	10:24:23 AM	84 min	T7790071	1/18/2027	492.00	484.00	1.63 %	Yes
phx42-3536	9/22/2022	A. Scarborough	Standard	0.30	08:59:25 AM	10:24:57 AM	84 min	CT20546	10/6/2026	9923.00	9812.00	1.12 %	Yes
phx42-3584	9/22/2022	A. Scarborough	Standard	0.30	08:59:25 AM	09:40:03 AM	39 min	CT-24077	7/18/2027	0.00	1.31	13.10 %	Yes
phx42-3584	9/22/2022	A. Scarborough	Standard	0.30	08:59:25 AM	09:41:45 AM	39 min	CT8327	1/19/2027	98.10	97.99	0.11 %	Yes
phx42-3584	9/22/2022	A. Scarborough	Standard	0.30	08:59:25 AM	09:43:26 AM	39 min	T7790071	1/18/2027	492.00	489.00	0.61 %	Yes
phx42-3584	9/22/2022	A. Scarborough	Standard	0.30	08:59:25 AM	09:45:07 AM	39 min	CT20546	10/6/2026	9923.00	9808.00	1.16 %	Yes
phx42-3590	9/22/2022	A. Scarborough	Standard	0.30	08:59:38 AM	09:39:29 AM	39 min	CT-24077	7/18/2027	0.00	0.82	8.20 %	Yes
phx42-3590	9/22/2022	A. Scarborough	Standard	0.30	08:59:38 AM	09:41:12 AM	39 min	CT8327	1/19/2027	98.10	97.73	0.38 %	Yes
phx42-3590	9/22/2022	A. Scarborough	Standard	0.30	08:59:38 AM	09:42:52 AM	39 min	T7790071	1/18/2027	492.00	488.00	0.81 %	Yes
phx42-3590	9/22/2022	A. Scarborough	Standard	0.30	08:59:38 AM	09:44:34 AM	39 min	CT20546	10/6/2026	9923.00	9878.00	0.45 %	Yes
phx42-3599	9/22/2022	A. Scarborough	Standard	0.00	12:05:20 PM	12:21:06 PM	16 min	CT-24077	7/18/2027	0.00	1.20	12.00 %	Yes
phx42-3599	9/22/2022	A. Scarborough	Standard	0.00	12:05:20 PM	12:21:41 PM	16 min	CT8327	1/19/2027	98.10	98.11	-0.01 %	Yes
phx42-3599	9/22/2022	A. Scarborough	Standard	0.00	12:05:20 PM	12:22:15 PM	16 min	T7790071	1/18/2027	492.00	491.00	0.20 %	Yes
phx42-3599	9/22/2022	A. Scarborough	Standard	0.00	12:05:20 PM	12:22:49 PM	16 min	CT20546	10/6/2026	9923.00	9804.00	1.20 %	Yes
phx42-5389	9/22/2022	S. Mort	Standard	0.30	10:16:25 AM	10:32:11 AM	16 min	CT-24077	7/18/2027	0.00	1.21	12.10 %	Yes
phx42-5389	9/22/2022	S. Mort	Standard	0.30	10:16:25 AM	10:32:45 AM	16 min	CT8327	1/19/2027	98.10	99.23	-1.15 %	Yes
phx42-5389	9/22/2022	S. Mort	Standard	0.30	10:16:25 AM	10:33:19 AM	16 min	T7790071	1/18/2027	492.00	486.00	1.22 %	Yes
phx42-5389	9/22/2022	S. Mort	Standard	0.30	10:16:25 AM	10:33:53 AM	16 min	CT20546	10/6/2026	9923.00	9834.00	0.90 %	Yes

Personnel have performed the required daily calibration, for all working Analyzers on site, to every calibration span gas.

Zero gas was used to purge after the initial calibration process and between all samples during the drift assessment and precision calibration processes.

Signature: _____	Date
Print: _____	
Field Supervisor	

Signature: _____	Date
Print: _____	
Site Administrator	

Signature: _____	Date
Print: _____	
Client Rep. (optional)	